



IP Human Papillomavirus [HPV] 14 high risk Viruses Real-time PCR Kit

[HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59,
66, 68]

Instructions for use

For use with: Agilent, Bio-Rad, Roche Lightcycler-96, Roche-Z480/Cobas-480, Applied Bio systems [ABI], Thermo-Piko-Real, Rotor gene 5/6plex, Alta-96, Cepheid Real time PCR machines.



IP500019-2



25x2

Contents

Intended use	5
Kit components	5
Storage	6
Material and Instruments required	6
Product use limitations	7
Product description	9
Specificity & Linear range	9
Analytical sensitivity	10
Note – DNA purification	10
Endogenous control system	11
Standards preparation	12
Detection protocol – reaction setup	13
Programming – Thermal profile & Dye selection	14
Reading the graph	15
Qualitative interpretation	16
Copies number calculation	16
Limitations	18
Quality control	19

Intended Use

The Immunoport HPV 14 high-risk virus Real-time PCR Kit is an in vitro nucleic acid amplification kit for the detection of HPV 14 high risk HPV virus specific DNA. Among 14 high risk HPV viruses, it detects and differentiates HPV 16 & 18 specific DNA.

Kit components

Components	Volume Per reaction	Number of vials	Volume Per vials
Probe PCR Master Mix	10µl	4	250µl
HPV 16 & 18 Primer Probe Mix [HPV 16 & 18 PP mix]	5µl	2	125µl
HPV 12 high risk viruses Primer Probe Mix [HPV 12 PP mix]	2.5µl	2	65µl
Endogenous control Primer Probe Mix [Endogenous PP mix]	2.5µl	2	65µl
Positive control mix [QS1]	10µl	2	250µl
Water, PCR grade		2	4ml

Storage

- The kit is shipped on gel ice. Upon arrival, all components should be stored in -20°C. They are stable until the expiration date stated on the label.
- Repeated thawing and freezing should be avoided, as this might affect the performance of the assay.
- If the reagents are to be used only intermittently, they should be frozen in aliquots. Storage at 2 to 8°C should not exceed a period of 5 hours.

Material and instruments required

- Real-time PCR instrument having FAM & HEX channels
- Automatic Nucleic acid extraction system or spin column based purification kit for the purification of nucleic acids
- Desktop centrifuge having 13000rpm or above with a rotor for 1.5/2 ml reaction tubes
- Centrifuge with a rotor for PCR strips/tubes and 96 well plates
- Optical cap qPCR tubes or strips or 96 wells
- Micro Pipettes (variables)
- Micro Pipette tips with filters (disposable)
- Powder-free gloves (disposable)

[Please ensure that all instruments used have been installed, calibrated, checked and maintained according to the manufacturer's instructions and recommendations.]

Product Use Limitations

- All reagents may exclusively be used in molecular diagnosis.
- The product is to be used by personnel specially instructed and trained in Molecular diagnosis.
- Strict compliance with the user manual is required for optimal PCR results.
- Attention should be paid to expiration dates printed on the box and labels of all components. Do not use expired components.
- Wear protective disposable powder-free gloves, a laboratory coat and eye protection when handling specimens and kit components.
- Avoid microbial and nuclease (DNase/RNase) contamination of the specimens and the components of the kit.
- Always use DNase/RNase-free disposable pipette tips with aerosol barriers.
- Use separated and segregated working areas for sample preparation, reaction setup and amplification/detection activities.
- The workflow in the laboratory should proceed in unidirectional manner. Always wear disposable gloves in each area and change them before entering a different area.
- Store positive and/or potentially positive material separated from all other components of the kit.
- Do not open the reaction tubes/plates post amplification, to avoid contamination with amplicons.

- Additional controls may be tested according to guidelines or requirements of local, state and/or federal regulations or accrediting organizations.
- Do not autoclave reaction tubes after the PCR, since this will not degrade the amplified nucleic acid and will bear the risk to contaminate the laboratory area.
- Discard sample and assay waste according to your local safety regulations.

Product description

Immunoport HPV 14 high-risk virus Real-time PCR Kit constitutes a ready-to-use system for the detection of HPV 14 high risk viruses' specific DNA using polymerase chain reaction (PCR). It contains reagents and enzymes for the specific amplification of the conserved region of the HPV genome, and for the direct detection of the specific amplicon in FAM & HEX channel. In addition, it contains an endogenous control [Human gene] amplification system to identify possible PCR inhibition and DNA purification efficiency. External positive control is supplied, which can be used as both qualitative and quantitative to determination the amount of viral load.

Specificity

HPV 14 high risk virus primer and probe have been designed for the specific and exclusive *in vitro* detection of HPV 14 high risk viruses [HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66, and 68]. The target sequence (L6/L7 gene) is highly conserved and has previously been shown to be a good genetic marker for HPV. The primers and probe sequences in this kit have 100% homology with a broad range of clinically relevant reference sequences based on a comprehensive bioinformatics analysis.

Dynamic linear range

The linear range was evaluated by analyzing a logarithmic dilution series of DNA concentrations ranging from 1.00E+09 to 1.00E+00 copies/ μ l. At least six replicates per dilution were analyzed. The linear range is 1.00E+09 to 1.00E+00 copies/ μ l.

Analytical Sensitivity

The analytical sensitivity is defined as the concentration of DNA molecules (copies/ μ l) that can be detected with a positivity rate of 95%. The analytical sensitivity was determined by analysis of dilution series of quantified HPV specific DNA from 0.001copies to 10copies/ μ l in triplicates. Under optimal PCR conditions, the analytical sensitivity is 0.25 copies per micro liter.

Note:**DNA Purification**

Purified DNA is the starting material for the Real-time PCR assay. The quality of the purified DNA has a profound impact on the performance of the entire test system. It has to be ensured that the purification system used for DNA purification is compatible with real-time PCR technology.

If you are using a spin column based sample preparation procedure having washing buffers containing ethanol, it is highly recommended to perform an additional centrifugation step for 5min at approximately 17000 x g (~ 13000 rpm), using a new collection tube, prior to the elution of the DNA.

Endogenous control

Human gene is given as endogenous control. It amplifies a single copy human gene from the test samples. A successful amplification indicates that test samples are properly collected and has its biological property.

The primer and probe present at PCR limiting concentrations which allows multiplexing with the target sequence primers. Amplification of the endogenous control template does not interfere with detection of the mutation even when present at low copy number. The endogenous control is detected through the HEX channel and gives a CT value of 21 +/-10.

Preparation of standard curve dilution series:

1. Pipette 90µl of PCR grade water into three 1.5ml micro centrifuge tubes and label as QS2 to QS4.
2. Pipette 10µl of Positive control-QS1 into tube QS2.
3. Vortex thoroughly and spin down briefly.
4. Change pipette tip and pipette 10µl from tube QS2 into tube QS3.
5. Vortex thoroughly and spin down briefly.
6. Repeat steps 4 and 5 to complete the dilution series.
7. Use 10µl per reaction.
8. Prepare fresh every time.

Standards	Copies per µl
QS-1	2500000
QS-2	250000
QS-3	25000
QS-4	2500

Detection Protocol

Things to do before starting

- Before use, all kit components need to be thawed completely, mixed by gently inverting and centrifuged briefly.
- Make sure that Positive and Negative control is included in every run.
- Include 0.5 reaction volume for pipetting error while calculating the volume for total number of reactions.

Components	Volume per reaction	
	HPV 16 & 18	HPV 12 & Endo
Probe PCR Master Mix	10µl	10µl
HPV 16 & 18 PP Mix	5µl	---
HPV 12 PP Mix	--	2.5µl
Endogenous PP Mix	--	2.5µl
Purified DNA	10µl	10µl
Final reaction volume	25µl	25µl

Negative Control setup [NTC]

Add 10µl of PCR grade water.

Qualitative Positive Control setup

Add 10µl of any one of the Positive controls [From QS1 to QS4]

Quantitative Positive controls setup [Suggested to use HPV 16/18 PP mix for Quantitative]

10µl of all Positive controls prepared from QS1 to QS4.

Centrifuge PCR vials briefly before placing into thermal cycler.
[Note: There should not be any bubbles in the reaction mix. Bubbles interfere with fluorescence detection.]

Programming Thermal cycler

Sample volume	25µl
Fluorescence Dyes	FAM & HEX
Passive reference	None
Ramping rate	Default

Thermal Profile

	Step	Time	Temp
	Taq enzyme activation / Hold	15min	95°C
45 cycles	Denaturation	20sec	95°C
	Annealing/Data collection*	20sec	56°C
	Extension	20sec	72°C

Data collection/Acquisition	Targets
FAM	HPV 16 & HPV 12
HEX	HPV 18 & Endogenous control

Reading the graph:**Step-1 – Endogenous control Validation**

Select the test samples alone for the endogenous control analysis. Select HEX dye and view the graph of endogenous control amplification. A successful amplification Ct value must be within Ct 21 +/- 10.

This range indicates NO PCR inhibition in the reaction. Any sample value goes beyond Ct 33 indicates that either sample has some issues in the purification or inhibiting PCR reaction.

Endogenous control will not get amplified in the negative and positive controls. Ignore a late noise HEX amplification graph in the NTC and Positive control wells.

Step-2 – FAM - Negative and Positive control validation

Select the NTC and Positive control [Qualitative] or Standards wells [Quantitative], select FAM channel, and view the graph of amplification.

The NTC must be flat with no Ct value. If required adjust the threshold value just above the NTC. The PC or Standards must be amplified as per their copy numbers.

NTC justifies NO contamination in the reagent as well as fine pipetting and its environment. PC justifies the reagents storage conditions and reaction parameters are as prescribed.

Step-3 –FAM - Test Sample status

In FAM channel, select test sample well one by one, analyze the graph/amplification.

Qualitative interpretation of results:

Test Sample HPV 16/18 /12types	Negative control	Positive control	Endogenous Control	Interpretation
Positive	Negative	Positive	Positive	HPV specific DNA detected
Negative	Negative	Positive	Positive	No HPV specific DNA Detected. Sample does not contain detectable amounts of HPV specific DNA.
Negative	Negative	Negative	Negative	Experiment fail
Positive	Positive	Positive	Positive	Experiment fail

Calculating copies per ml

Input the machine indicated copy number into the following formula

$$\text{Result (copies/}\mu\text{l)} \times \text{Elution Volume (}\mu\text{l)}$$

$$\text{Result (copies/ml)} = \frac{\text{-----}}{\text{Sample Volume (ml)}}$$

Note:

Elution volume: must be typed in micro liter format, example 30 μ l, 60 μ l or 100 μ l. **Sample volume:** must be typed in milliliter format, example 0.2ml or 0.5ml

Limitations

Good laboratory practice is essential for proper performance of this assay. Strict compliance with the instructions for use is required for optimal results.

Analysts should be trained and familiar with testing procedures and interpretation of results prior to performing the assay.

A false negative result may occur if inadequate numbers of organisms are present in the sample due to improper collection, transport or handling. Appropriate specimen collection, transport, storage and processing procedures are required for the optimal performance of this test.

Extreme care should be taken to preserve the purity of the components of the kit and reaction setups. All reagents should be closely monitored for impurity and contamination. Any suspicious reagents should be discarded.

The presence of PCR inhibitors may cause under quantification, false negative or invalid results.

Potential mutations within the target regions of the pathogen's genome covered by the primers and/or probes used in the kit may result in under quantification and/or failure to detect.

As with any diagnostic test, the Immunoport HPV 14 high risk virus Real-time PCR results need to be interpreted in consideration of all clinical and laboratory findings.

Quality Control

In accordance with the Immunoport Biomolecules in house Quality Management System, each lot of Immunoport HPV 14 high risk virus Real-time PCR kit is tested against predetermined specifications to ensure consistent product quality.

India Contact:

Life Technologies (India) Pvt. Ltd.

Ph: +91-11-42208000, Mobile: +91-9810521400

Email: customerservice@lifetechindia.com Website: www.lifetechindia.com