

## Aflatoxin M1 (AFM1) ELISA test kit

### INSTRUCTION MANUAL

FOR PRODUCT No: LT23001AYSLS



### Arsh Biotech Pvt. Ltd.

308, Aggarwal City Mall, Road No.44,  
Pitampura, Delhi-110034, India  
Mobile: +91-98105-21400 | Fax: +91-11-42208444  
info@arshbiotech.com

### PRINCIPLE

This test kit is based on the indirect competitive enzyme immunoassay for the detection of Aflatoxins M1. The coupling antigen is pre-coated on the micro-well stripes. The Aflatoxins M1 in the sample and the coupling antigens pre-coated on the micro-well stripes compete for the anti- Aflatoxins M1 antibodies. After the addition of the enzyme conjugate, the TMB substrate is added for coloration. The optical density (OD) value of the sample has a negative correlation with the Aflatoxins M1 in the sample. This value is compared to the standard curve and the Aflatoxins M1 residues is subsequently obtained.

### MATERIALS SUPPLIED WITH THIS KIT

1	Micro-well strips	12 strips with 8 removable wells each	6	Substrate B	7 ml, black cap
2	6× standard solution (1 ml each)	0ppb, 0.03ppb, 0.09ppb, 0.27ppb, 0.81ppb, 2.43ppb	7	Stop solution	7 ml, yellow cap
3	Enzyme conjugate	7 ml, red cap	8	20× concentrated wash	40 ml, white cap
4	Antibody working solution	7 ml, blue cap	9	10× concentrated redissolving solution	50 ml, transparent cap
5	Substrate A	7 ml, white cap			

### MATERIALS REQUIRED BUT NOT PROVIDED

- 1) Micropipettors and disposable tips: 0.5µl-10µl, 10µl-100µl, 100µl-1000µl.
- 2) 37 °C Incubator.
- 3) Measuring cylinder: 500 ml.
- 4) 96 wells microplate reader.
- 5) Distilled/De-ionized water.
- 6) Microplate Washer.

MANUAL VERSION 1.01

## TECHNICAL SPECIFICATIONS

**Sensitivity:** 0.03 ppb

**Incubator temperature:** 25°C

**Incubator time:** 15min-30min

### Detection limit:

Milk – 0.1ppb, Milk powder, Yogurt – 0.3ppb

### Cross-reaction rate:

Aflatoxin M1... ..1.00%

Aflatoxin B1... ..12.9%

Aflatoxin B2... ..1.4%

Aflatoxin G1... ..1.9%

Aflatoxin G2... ..0.3%

### Recovery rate:

Milk - 90±25%

Milk powder, Yogurt - 95±30%

## SAMPLE PRE-TREATMENT

*Instructions* (The following points must be dealt with before the pre-treatment)

- 1) Only the disposable tips can be used for the experiments and the tips must be changed when used for absorbing different reagents;
- 2) Before the experiment, each experimental utensil must be clean and should be re-cleaned if necessary, in order to avoid the contamination that interferes with the experimental results.

### Solution preparation before sample pre-treatment:

- 1) **Sample redissolving solution**

Use 1 part of 10X concentrated re-dissolving solution and dissolve with 9 parts of deionized water to obtain the ready to use sample re-dissolving solution.

## SAMPLES PREPARATION

### Preparation of raw milk and finished milk samples

- 1) Take raw milk and finished milk samples, put them at room temperature, test them directly when samples return back to room temperature.

- 2) Take 50 µl to test

Dilution factor: 1

### Preparation of milk powder sample

- 1) Take 1.0±0.05g milk powder sample into a 50ml centrifuge tube; add 10ml Sample re-dissolving solution, shake thoroughly for 3min, centrifuge at above 4000r/min at 20°C for 10 min;

- 2) Take 50µl supernatant to test.

Dilution factor: 10

### Preparation of Yogurt sample

- 1) Take 1ml yogurt sample, dilute with Sample redissolving solution at 1:9 (100ul yogurt + 900ul Sample redissolving solution), mix for 30s;

- 2) Take 50 µl to test

Dilution factor: 10

## ELISA PROCEDURES

### Instructions

1. Bring ELISA reagents to room temperature (20 - 25 °C) before use.

2. Put ELISA reagents back to 2-8 °C immediately after use.
3. The ELISA reproducibility in the analysis process is largely depends on the consistency of the washing plate, the correct washing plate operation is the point of determination ELISA program.
4. In all process of constant temperature incubation, avoid light exposure, seal the microplate with the cover membrane.

### PROTOCOL

1. Bring test kit to the room temperature (20-25 °C) for at least 30 min, note that each reagent must be shaken evenly before use;
2. Put the required micro-well strips into plate frames. Re-seal the unused microplate, stored at 2-8 °C, not frozen.
3. Solution preparation: take the 40ml 20× concentrated washing buffer, dissolve with deionized water at 1:19 (1 part 20× concentrated washing buffer + 19 parts deionized water), or prepare as quantity needed.
4. Numbering: number the micro-wells according to samples and standard solution; each sample and standard solution should be performed in duplicate; record their positions.
5. Add standard/sample: Add 50 µl of the sample or the standard solution into separate duplicate wells, then add enzyme conjugate, 50 µl/well; then antibody working solution, 50 µl/well. Mix gently by shaking the plate manually, seal the microplate with the cover membrane, incubate at 25 °C for 30 min in the dark.

6. Wash microplate: Carefully open the cover membrane, pour liquid out of microwell; add 250 µl/well of washing buffer, wash fully for 4-5 times, 15-30 s each time, then take out and flap to dry with absorbent paper (Use

unused spear to pierce bubble after dry).

7. Coloration: add 50 µl of substrate A solution then 50 µl B solution into each well. Mix gently by shaking the plate manually, and incubate at 25 °C for 15min in the dark for coloration.
8. Determination: add 50 µl of the stop solution into each well. Mix gently by shaking the plate manually. Set the wavelength of the microplate reader at 450 nm to determine the OD value of every well (Recommend to read the OD value at the dual-wavelength 450/630 nm within 5 min).

### RESULTS JUDGEMENT

There are two methods to judge the results; the first one is the rough judgment, while the second is the quantitative determination. Note that the OD value of the sample has a negative correlation with the Aflatoxin M1 in the sample

### QUALITATIVE DETERMINATION

The concentration range (ppb) can be obtained by compared the average absorbance value with standards. Suppose absorbance value of Sample One is 0.3, Sample Two is 1.0, and the standards are: 0ppb of 2.243, 0.03ppb of 1.816, 0.09ppb of 1.415, 0.27ppb of 0.74, 0.81ppb of 0.313, 2.43ppb of 0.155. Then the concentration of the sample one is in the range of 0.81ppb - 2.43ppb. Sample Two is 0.09ppb-0.27ppb. The concentration range of aflatoxin M1 in the samples can be obtained by multiplied by the corresponding dilution of the sample.

### QUANTITATIVE ANALYSIS

In order to calculate the concentration of samples, a standard curve should be

made. Before standard curve is made, the concept of % absorbance should be known.

**Calculation of % absorbance:**

Percentage of absorbance value =  $B/B_0 \times 100\%$

B—the average OD value of the sample or the standard solution

B<sub>0</sub>—the average OD value of the 0 ng/ml standard solution

The zero standard is thus made equal to 100 % and the absorbance values are quoted in percentages. The values calculated for the standards are entered in a system of coordinates on semilogarithmic graph paper against the aflatoxin B1 concentration [ng/l]. The aflatoxin B1 concentration in ng/l (ppb) corresponding to the absorbance of each sample can be read from the calibration curve.

Special software for result analysis of ELISA will facilitate double or multiple determinations. If you need, please call to request.

**PRECAUTIONS**

1. The room temperature below 25 °C or the temperature of the reagents and the samples being not returned to the room temperature (20-25 °C) will lead to a lower standard OD value.
2. Dryness of the microplate in the washing process will be accompanied by the situations including the non-linear standard curves and the undesirable reproducibility; So continue to next step immediately after washing.
3. Mix evenly before adding any reagents.
4. The stop solution is the 2 M sulfuric acid solution, avoid contacting with the skin.

5. Do not use the kit exceeding its expiry date. The use of diluted or adulterated reagents from the kits will lead to the changes in the sensitivity and the detecting OD values. Do not exchange the reagents from the kits of different lot numbers to use.

6. Storage: store at 2-8 °C, not frozen. Put the unused microplate into an auto-sealing bag to re-seal it. The standard substance and the colourless color former is light sensitive, and thus they cannot be directly exposed to the light.

7. Discard the colouration solution with any color that indicates the degeneration of this solution. The detecting value (450/630nm) of the 0 standard solution (0 ppb) of less than 0.5 (A450nm<0.5) indicates its degeneration.

8. The optimum reaction temperature is 25 °C, and too high or too low temperatures will result in the changes in the detecting sensitivity and OD values.

**STORAGE AND EXPIRY DATE**

**Storage: store at 2-8 °C, not frozen.**

**Expiry date: 12 months; date of production is on box.**

**Note: If the Vacuum package of microplate has leakage, it is still valid to use, do not affect the test result, be relax to use.**